

30 Jun '48

MDDC - 1051

UNITED STATES
ATOMIC ENERGY COMMISSION
OAK RIDGE
TENNESSEE

184" CYCLOTRON
VERTICAL D. C. ELECTROSTATIC DEFLECTOR

by

Duane Sewell


Published for use within the Atomic Energy Commission. Inquiries for additional copies and any questions regarding reproduction by recipients of this document may be referred to the Technical Information Division, Atomic Energy Commission, P. O. Box E, Oak Ridge, Tennessee.

Inasmuch as a declassified document may differ materially from the original classified document by reason of deletions necessary to accomplish declassification, this copy does not constitute authority for declassification of classified copies of a similar document which may bear the same title and authors.

Date of Manuscript: February 18, 1947

Document Declassified: June 10, 1947

This document consists of 2 pages.



184" CYCLOTRON VERTICAL D. C. ELECTROSTATIC DEFLECTOR

By Duane Sewell

Experiment done by: J. Vale, L. Houser, R. Watt, W. Chupp,
A. Reyenga, and W. Grimshaw

INTRODUCTION

It was found from previous experiments that the beam was rapidly spreading vertically at approximately the 81" radius. It was thought that if a vertical deflector was placed below the median plane at this 81" radius, the beam could then expand into the deflecting channel. Once in the deflecting channel it could then be deflected rapidly into a region in which a magnetic shield (or deflector) could be used to bring the beam out of the cyclotron.

RESULTS

1. Approximately 60 kv could be held on the high voltage electrode of this deflector.
2. On one occasion it looked as if an ionization chamber mounted on the probe was detecting a small deflection of the beam. However, this could not be verified on subsequent runs.
3. The best detection of the beam deflection was made by mounting X-ray film on the probe and exposing it to both the undeflected and deflected beam.
4. There was considerable sparking between the HV electrode and the .002" copper foil. The copper foil was warped badly when the deflector was removed; this had apparently been caused by the local heating from the sparking. The 0.002" copper foil supported in this manner is not suitable for this job.
5. The copper foil was very radioactive after a short run.

EXPERIMENTAL SET UP:






Figure 1

184" CYCLOTRON
VERTICAL D. C. ELECTROSTATIC DEFLECTOR

EXPERIMENTAL SET UP:



VERTICAL D.C. INSULATOR BY VICTOR



UNIVERSITY OF FLORIDA



3 1262 08909 7652